

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003290.D
 Acq On : 10 Jul 2018 17:55
 Operator : JC/MD
 Sample : PB110944ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110944ZHE#04

Quant Time: Jul 11 01:11:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	354936	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	336753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	190787	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	161807	40.45	ug/l	0.00
Spiked Amount			Recovery	=	80.90%	
35) Dibromofluoromethane	5.51	113	137242	35.99	ug/l	0.00
Spiked Amount			Recovery	=	71.98%	
50) Toluene-d8	8.72	98	500765	36.97	ug/l	0.00
Spiked Amount			Recovery	=	73.94%	
62) 4-Bromofluorobenzene	11.14	95	171883	34.34	ug/l	0.00
Spiked Amount			Recovery	=	68.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	1772	6.10	ug/l #	95
16) Acetone	2.45	43	26813	18.88	ug/l	91
18) Methyl Acetate	2.78	43	23013	5.74	ug/l	91
20) Methylene Chloride	2.86	84	7700	1.97	ug/l	95
25) 2-Butanone	4.72	43	4889	1.95	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003290.D
Acq On : 10 Jul 2018 17:55
Operator : JC/MD
Sample : PB110944ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110944ZHE#04

Quant Time: Jul 11 01:11:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

