

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003300.D
 Acq On : 10 Jul 2018 22:23
 Operator : JC/MD
 Sample : J3892-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGEBLANK-SAMPLEMANAG

Quant Time: Jul 11 01:32:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252205	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	348682	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	335883	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184893	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	158729	39.99	ug/l	0.00
Spiked Amount			Recovery	=	79.98%	
35) Dibromofluoromethane	5.51	113	136721	36.50	ug/l	0.00
Spiked Amount			Recovery	=	73.00%	
50) Toluene-d8	8.72	98	499549	37.54	ug/l	0.00
Spiked Amount			Recovery	=	75.08%	
62) 4-Bromofluorobenzene	11.14	95	173560	35.30	ug/l	0.00
Spiked Amount			Recovery	=	70.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003300.D
Acq On : 10 Jul 2018 22:23
Operator : JC/MD
Sample : J3892-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SAMPLEMANAG

Quant Time: Jul 11 01:32:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

