

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003308.D
 Acq On : 11 Jul 2018 02:50
 Operator : JC/MD
 Sample : J3914-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B3

Quant Time: Jul 11 05:18:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	259216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	358627	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	340209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	191061	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	164451	40.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.62%	
35) Dibromofluoromethane	5.51	113	140150	36.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.76%	
50) Toluene-d8	8.72	98	508800	37.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.36%	
62) 4-Bromofluorobenzene	11.14	95	174062	34.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.84%	
Target Compounds						
16) Acetone	2.45	43	35694	24.65	ug/l	99
18) Methyl Acetate	2.78	43	19579	4.79	ug/l	96
25) 2-Butanone	4.71	43	15234	5.94	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003308.D
Acq On : 11 Jul 2018 02:50
Operator : JC/MD
Sample : J3914-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B3

Quant Time: Jul 11 05:18:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

