

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003315.D
 Acq On : 11 Jul 2018 05:57
 Operator : JC/MD
 Sample : J3908-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-10-S

Quant Time: Jul 11 06:43:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	360754	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	340515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	192619	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	159577	38.05	ug/l	0.00
Spiked Amount			Recovery	=	76.10%	
35) Dibromofluoromethane	5.51	113	141221	36.44	ug/l	0.00
Spiked Amount			Recovery	=	72.88%	
50) Toluene-d8	8.72	98	504585	36.65	ug/l	0.00
Spiked Amount			Recovery	=	73.30%	
62) 4-Bromofluorobenzene	11.14	95	174509	34.31	ug/l	0.00
Spiked Amount			Recovery	=	68.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	1125	5.93	ug/l #	88
16) Acetone	2.45	43	20856	14.01	ug/l	99
18) Methyl Acetate	2.78	43	18061	4.30	ug/l	94
20) Methylene Chloride	2.86	84	38239	11.99	ug/l	94
25) 2-Butanone	4.72	43	4837	1.84	ug/l #	90
95) Naphthalene	13.84	128	78022	5.07	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003315.D
Acq On : 11 Jul 2018 05:57
Operator : JC/MD
Sample : J3908-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-10-S

Quant Time: Jul 11 06:43:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

