

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071018\
 Data File : VX003298.D
 Acq On : 10 Jul 2018 21:29
 Operator : JC/MD
 Sample : J3892-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE BLANK-SOILREF-5

Quant Time: Jul 11 01:28:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	247342	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	341334	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	318877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	183507	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	157424	40.44	ug/l	0.00
Spiked Amount			Recovery	=	80.88%	
35) Dibromofluoromethane	5.51	113	134454	36.67	ug/l	0.00
Spiked Amount			Recovery	=	73.34%	
50) Toluene-d8	8.72	98	481952	37.00	ug/l	0.00
Spiked Amount			Recovery	=	74.00%	
62) 4-Bromofluorobenzene	11.14	95	165206	34.33	ug/l	0.00
Spiked Amount			Recovery	=	68.66%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071018\
Data File : VX003298.D
Acq On : 10 Jul 2018 21:29
Operator : JC/MD
Sample : J3892-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SOILREF-5

Quant Time: Jul 11 01:28:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.01 min

Lab File: VX003298.D

Acq: 10 Jul 2018 21:29

Instrument :

MSVOA_X

ClientSampleId :


STORAGEBLANK-SOILREF-5

Tot Ion:168 Resp: 247342

Ion Ratio Lower Upper

168 100

99 47.5 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 40.44 ug/l

RT: 6.07 min Scan# 900

Delta R.T. -0.00 min

Lab File: VX003298.D


Acq: 10 Jul 2018 21:29

Tgt Ion: 65 Resp: 157424

Ion Ratio Lower Upper

65 100

67 52.2 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

60000


40000


20000

0

Time-->

6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX003298.D

Acq: 10 Jul 2018 21:29

Instrument :

MSVOA_X

ClientSampleId :

STORAGEBLANK-SOILREF-5


Tot Ion:114 Resp: 341334

Ion Ratio Lower Upper

114 100

63 19.7 0.0 41.4


88 15.3 0.0 30.0





Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



Time-->



#35

Dibromofluoromethane

Concen: 36.67 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX003298.D

Acq: 10 Jul 2018 21:29


Tgt Ion:113 Resp: 134454

Ion Ratio Lower Upper

113 100

111 101.1 81.4 122.0


192 24.6 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V


Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



Time-->









#50
Toluene-d8
Concen: 37.00 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. -0.00 min
Lab File: VX003298.D
Acq: 10 Jul 2018 21:29




Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SOILREF-5


Tot Ion: 98 Resp: 481952
Ion Ratio Lower Upper
98 100
100 64.0 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 34.33 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. -0.00 min
Lab File: VX003298.D
Acq: 10 Jul 2018 21:29

Tgt Ion: 95 Resp: 165206
Ion Ratio Lower Upper
95 100
174 100.0 0.0 187.4
176 96.7 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX003298.D
Acq: 10 Jul 2018 21:29

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-SOILREF-5

Tot Ion:117 Resp: 318877
Ion Ratio Lower Upper
117 100
82 52.8 45.0 67.4
119 31.9 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1886
Delta R.T. 0.01 min
Lab File: VX003298.D
Acq: 10 Jul 2018 21:29

Tgt Ion:152 Resp: 183507
Ion Ratio Lower Upper
152 100
115 52.5 40.1 120.2
150 153.7 0.0 347.2

