

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX071019\
 Data File : VX010774.D
 Acq On : 10 Jul 2019 14:02
 Operator : JC/SP
 Sample : K3732-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030LC068-SW-190709

Quant Time: Jul 11 01:59:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	210130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	330924	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	304631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	142847	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	106382	54.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.00%	
35) Dibromofluoromethane	5.50	113	101405	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
50) Toluene-d8	8.71	98	397071	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
62) 4-Bromofluorobenzene	11.13	95	133007	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
Target Compounds						
16) Acetone	2.45	43	3988	4.039	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX071019\
Data File : VX010774.D
Acq On : 10 Jul 2019 14:02
Operator : JC/SP
Sample : K3732-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY030LC068-SW-190709

Quant Time: Jul 11 01:59:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

