

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX071019\
 Data File : VX010783.D
 Acq On : 10 Jul 2019 17:33
 Operator : JC/SP
 Sample : K3738-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Jul 11 02:36:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	207043	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	325796	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	302362	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137721	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107158	55.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.40%	
35) Dibromofluoromethane	5.50	113	100892	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
50) Toluene-d8	8.71	98	388448	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	11.13	95	131888	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
Target Compounds						
4) Vinyl Chloride	1.41	62	2938	1.600	ug/l	93
8) Diethyl Ether	2.19	74	27446	26.298	ug/l	99
16) Acetone	2.45	43	3950	4.060	ug/l #	88
22) Diisopropyl ether	3.86	45	7613	1.429	ug/l #	75
27) cis-1,2-Dichloroethene	4.60	96	23613	10.587	ug/l	92
42) 1,2-Dichloroethane	6.21	62	1941	0.791	ug/l	84
44) Trichloroethene	7.21	130	15171	6.286	ug/l	97
64) Tetrachloroethene	9.34	164	1622	0.663	ug/l #	88
65) Chlorobenzene	10.14	112	4563	0.743	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX071019\
Data File : VX010783.D
Acq On : 10 Jul 2019 17:33
Operator : JC/SP
Sample : K3738-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Jul 11 02:36:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

