

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX071019\
 Data File : VX010787.D
 Acq On : 10 Jul 2019 19:06
 Operator : JC/SP
 Sample : PB121247TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121247TB

Quant Time: Jul 11 03:14:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	213350	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	340133	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	324244	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149711	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110765	55.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.74%	
35) Dibromofluoromethane	5.50	113	106279	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
50) Toluene-d8	8.71	98	415904	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
62) 4-Bromofluorobenzene	11.13	95	140913	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
Target Compounds						
16) Acetone	2.45	43	13052	13.019	ug/l	98
20) Methylene Chloride	2.86	84	3971	1.925	ug/l	91
43) Isopropyl Acetate	6.45	43	13310	2.893	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX071019\
Data File : VX010787.D
Acq On : 10 Jul 2019 19:06
Operator : JC/SP
Sample : PB121247TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121247TB

Quant Time: Jul 11 03:14:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

