

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071020\
 Data File : VX017327.D
 Acq On : 10 Jul 2020 11:52
 Operator : JC/SP
 Sample : VX0710WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0710WBL02

Quant Time: Jul 11 01:17:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	234634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	384768	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	394837	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	198954	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	154091	53.27	ug/l	0.00
Spiked Amount						
			Recovery	=	106.54%	
35) Dibromofluoromethane	5.46	113	130525	53.63	ug/l	0.00
Spiked Amount						
			Recovery	=	107.26%	
50) Toluene-d8	8.70	98	467462	51.94	ug/l	0.00
Spiked Amount						
			Recovery	=	103.88%	
62) 4-Bromofluorobenzene	11.12	95	187955	53.42	ug/l	0.00
Spiked Amount						
			Recovery	=	106.84%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071020\
Data File : VX017327.D
Acq On : 10 Jul 2020 11:52
Operator : JC/SP
Sample : VX0710WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0710WBL02

Quant Time: Jul 11 01:17:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

