

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071020\
 Data File : VX017336.D
 Acq On : 10 Jul 2020 15:33
 Operator : JC/SP
 Sample : PB130081ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130081ZHE#02

Quant Time: Jul 11 01:32:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	273385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	455566	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	479078	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	255348	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	171075	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
35) Dibromofluoromethane	5.47	113	147778	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	8.70	98	560713	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
62) 4-Bromofluorobenzene	11.12	95	242348	58.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.36%	
Target Compounds						
16) Acetone	2.42	43	26718	21.191	ug/l	93
18) Methyl Acetate	2.75	43	3800	1.155	ug/l	98
20) Methylene Chloride	2.84	84	13921	4.096	ug/l	96
43) Isopropyl Acetate	6.42	43	86857	11.965	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071020\
Data File : VX017336.D
Acq On : 10 Jul 2020 15:33
Operator : JC/SP
Sample : PB130081ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130081ZHE#02

Quant Time: Jul 11 01:32:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

