

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071020\
 Data File : VX017338.D
 Acq On : 10 Jul 2020 16:19
 Operator : JC/SP
 Sample : PB130081TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130081TB

Quant Time: Jul 11 01:35:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	255315	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	421963	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	465324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	240933	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	164082	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
35) Dibromofluoromethane	5.46	113	139939	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
50) Toluene-d8	8.70	98	532468	53.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.90%	
62) 4-Bromofluorobenzene	11.12	95	230341	59.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.40%	
Target Compounds						
16) Acetone	2.42	43	33154	28.157	ug/l	100
18) Methyl Acetate	2.75	43	3408	1.109	ug/l	99
20) Methylene Chloride	2.84	84	17921	5.947	ug/l	95
43) Isopropyl Acetate	6.42	43	75339	11.205	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071020\
Data File : VX017338.D
Acq On : 10 Jul 2020 16:19
Operator : JC/SP
Sample : PB130081TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130081TB

Quant Time: Jul 11 01:35:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

