

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071020\
 Data File : VX017339.D
 Acq On : 10 Jul 2020 16:41
 Operator : JC/SP
 Sample : L3184-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EO-02-070920

Quant Time: Jul 11 01:40:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	256446	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	428498	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	467229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	245120	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	163296	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
35) Dibromofluoromethane	5.47	113	139760	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
50) Toluene-d8	8.70	98	535609	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
62) 4-Bromofluorobenzene	11.12	95	230232	58.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.52%	
Target Compounds						
16) Acetone	2.42	43	68196	57.663	ug/l	96
20) Methylene Chloride	2.84	84	17115	5.615	ug/l	95
43) Isopropyl Acetate	6.42	43	46500	6.810	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071020\
Data File : VX017339.D
Acq On : 10 Jul 2020 16:41
Operator : JC/SP
Sample : L3184-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EO-02-070920

Quant Time: Jul 11 01:40:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

