

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071020\
 Data File : VX017342.D
 Acq On : 10 Jul 2020 17:50
 Operator : JC/SP
 Sample : L3259-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T-2-DRUM

Quant Time: Jul 11 01:47:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	253696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	428100	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	457865	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	241924	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	164419	52.57	ug/l	0.00
Spiked Amount			Recovery	=	105.14%	
35) Dibromofluoromethane	5.46	113	139998	51.70	ug/l	0.00
Spiked Amount			Recovery	=	103.40%	
50) Toluene-d8	8.70	98	530292	52.96	ug/l	0.00
Spiked Amount			Recovery	=	105.92%	
62) 4-Bromofluorobenzene	11.12	95	223475	57.09	ug/l	0.00
Spiked Amount			Recovery	=	114.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	15119	25.340	ug/l	98
16) Acetone	2.42	43	100717	86.083	ug/l	95
18) Methyl Acetate	2.75	43	3643	1.193	ug/l	94
20) Methylene Chloride	2.84	84	36586	13.056	ug/l	96
43) Isopropyl Acetate	6.42	43	61730	9.049	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071020\
Data File : VX017342.D
Acq On : 10 Jul 2020 17:50
Operator : JC/SP
Sample : L3259-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T-2-DRUM

Quant Time: Jul 11 01:47:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

