

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071020\
 Data File : VX017343.D
 Acq On : 10 Jul 2020 18:13
 Operator : JC/SP
 Sample : L3251-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BUR-DRUM-COMP

Quant Time: Jul 11 01:50:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	263292	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	437312	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	476618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	258841	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	167056	51.47	ug/l	0.00
Spiked Amount			Recovery	=	102.94%	
35) Dibromofluoromethane	5.46	113	142463	51.50	ug/l	0.00
Spiked Amount			Recovery	=	103.00%	
50) Toluene-d8	8.70	98	545433	53.32	ug/l	0.00
Spiked Amount			Recovery	=	106.64%	
62) 4-Bromofluorobenzene	11.12	95	244591	61.17	ug/l	0.00
Spiked Amount			Recovery	=	122.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	11577	18.696	ug/l	97
16) Acetone	2.42	43	423394	348.689	ug/l	97
18) Methyl Acetate	2.75	43	3543	1.118	ug/l	93
20) Methylene Chloride	2.84	84	24036	7.973	ug/l	98
25) 2-Butanone	4.64	43	49539	24.485	ug/l	100
43) Isopropyl Acetate	6.42	43	66739	9.577	ug/l	99
52) Toluene	8.76	92	83693	11.038	ug/l	98
68) m/p-Xylenes	10.34	106	9805	1.554	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071020\
Data File : VX017343.D
Acq On : 10 Jul 2020 18:13
Operator : JC/SP
Sample : L3251-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BUR-DRUM-COMP

Quant Time: Jul 11 01:50:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

