

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
 Data File : VX036475.D
 Acq On : 10 Jul 2023 11:42
 Operator : JC/MD
 Sample : VX0710MBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0710MBL01

Quant Time: Jul 11 01:25:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	191670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	325700	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140615	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	125429	45.653	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.300%
35) Dibromofluoromethane	5.385	113	104420	48.622	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.240%
50) Toluene-d8	8.647	98	398174	49.640	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.280%
62) 4-Bromofluorobenzene	11.079	95	143656	50.846	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
Data File : VX036475.D
Acq On : 10 Jul 2023 11:42
Operator : JC/MD
Sample : VX0710MBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0710MBL01

Quant Time: Jul 11 01:25:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

