

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
 Data File : VX036480.D
 Acq On : 10 Jul 2023 14:10
 Operator : JC/MD
 Sample : PB154051ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB154051ZHE#02

Quant Time: Jul 11 01:28:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	178885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	310291	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131840	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	119285	46.520	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.040%		
35) Dibromofluoromethane	5.385	113	99278	48.524	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.040%		
50) Toluene-d8	8.647	98	379842	49.707	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.420%		
62) 4-Bromofluorobenzene	11.079	95	137371	51.036	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	102.080%		
Target Compounds						
16) Acetone	2.380	43	172200	157.116	ug/l	96
20) Methylene Chloride	2.788	84	8934	3.786	ug/l	94
25) 2-Butanone	4.556	43	98095	62.293	ug/l	96
37) Ethyl Acetate	4.715	43	41283	13.281	ug/l	99
43) Isopropyl Acetate	6.336	43	174840	35.259	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	439691	137.630	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
Data File : VX036480.D
Acq On : 10 Jul 2023 14:10
Operator : JC/MD
Sample : PB154051ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB154051ZHE#02

Quant Time: Jul 11 01:28:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

