

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
 Data File : VX036482.D
 Acq On : 10 Jul 2023 14:57
 Operator : JC/MD
 Sample : PB154051ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB154051ZHE#04

Quant Time: Jul 11 01:29:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	177179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	301867	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	280489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	115639	45.532	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.060%
35) Dibromofluoromethane	5.385	113	94825	47.641	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.280%
50) Toluene-d8	8.647	98	375724	50.540	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.080%
62) 4-Bromofluorobenzene	11.079	95	132574	50.628	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.260%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	165853	152.782	ug/l	92
20) Methylene Chloride	2.788	84	11270	4.822	ug/l	90
25) 2-Butanone	4.556	43	96247	61.708	ug/l	97
37) Ethyl Acetate	4.721	43	39623	13.103	ug/l	98
43) Isopropyl Acetate	6.336	43	174456	36.164	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	431903	138.965	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
Data File : VX036482.D
Acq On : 10 Jul 2023 14:57
Operator : JC/MD
Sample : PB154051ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB154051ZHE#04

Quant Time: Jul 11 01:29:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

