

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
 Data File : VX036490.D
 Acq On : 10 Jul 2023 18:06
 Operator : JC/MD
 Sample : 03555-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-347

Quant Time: Jul 11 01:32:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	172432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	294097	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	275842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115070	46.555	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.120%
35) Dibromofluoromethane	5.385	113	92987	47.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.900%
50) Toluene-d8	8.647	98	363241	50.151	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.300%
62) 4-Bromofluorobenzene	11.079	95	127717	50.062	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.120%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	72242	68.380	ug/l	94
20) Methylene Chloride	2.788	84	6548	2.879	ug/l	93
25) 2-Butanone	4.556	43	51162	33.705	ug/l	99
37) Ethyl Acetate	4.721	43	10108	3.431	ug/l	94
43) Isopropyl Acetate	6.336	43	119474	25.421	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	212781	70.271	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
Data File : VX036490.D
Acq On : 10 Jul 2023 18:06
Operator : JC/MD
Sample : 03555-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-347

Quant Time: Jul 11 01:32:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

