

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
 Data File : VX036494.D
 Acq On : 10 Jul 2023 19:41
 Operator : JC/MD
 Sample : 03555-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-240

Quant Time: Jul 11 01:33:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	177444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	311413	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	120032	47.191	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.380%
35) Dibromofluoromethane	5.385	113	96905	47.193	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	8.647	98	384723	50.164	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.320%
62) 4-Bromofluorobenzene	11.079	95	136964	50.701	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.400%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	70107	64.485	ug/l #	89
20) Methylene Chloride	2.788	84	8598	3.673	ug/l	86
25) 2-Butanone	4.556	43	51492	32.964	ug/l	90
37) Ethyl Acetate	4.715	43	11737	3.762	ug/l	98
43) Isopropyl Acetate	6.342	43	124946	25.107	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	224776	70.105	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
Data File : VX036494.D
Acq On : 10 Jul 2023 19:41
Operator : JC/MD
Sample : 03555-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-240

Quant Time: Jul 11 01:33:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

