



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071023\
Data File : VX036473.D
Acq On : 10 Jul 2023 10:16
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Title : SW846 8260
Last Update : Fri Jun 23 15:03:28 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	12622	PASS
75	95	30	60	50.0	32317	PASS
95	95	100	100	100.0	64619	PASS
96	95	5	9	6.9	4432	PASS
173	174	0.00	2	1.0	509	PASS
174	95	50	100	75.5	48779	PASS
175	174	5	9	7.9	3868	PASS
176	174	95	101	95.5	46589	PASS
177	176	5	9	6.6	3096	PASS