

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042257.D
 Acq On : 10 Jul 2024 11:10
 Operator : JC/MD
 Sample : VX0710WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0710WBL01

Quant Time: Jul 11 00:52:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	170720	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	265474	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91866	46.665	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.340%
35) Dibromofluoromethane	5.385	113	88321	46.888	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.647	98	300413	45.397	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.800%#
62) 4-Bromofluorobenzene	11.079	95	111077	45.535	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
Data File : VX042257.D
Acq On : 10 Jul 2024 11:10
Operator : JC/MD
Sample : VX0710WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0710WBL01

Quant Time: Jul 11 00:52:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

