

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042260.D
 Acq On : 10 Jul 2024 12:26
 Operator : JC/MD
 Sample : PB161974ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161974ZHE#01

Quant Time: Jul 11 00:54:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	139666	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	232617	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	212598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84714	52.600	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.200%
35) Dibromofluoromethane	5.391	113	81796	49.557	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.120%
50) Toluene-d8	8.647	98	249144	42.967	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	85.940%#
62) 4-Bromofluorobenzene	11.079	95	93947	43.953	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.900%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	47706	67.459	ug/l	# 89
18) Methyl Acetate	2.709	43	1978	1.025	ug/l	99
20) Methylene Chloride	2.788	84	8250	5.328	ug/l	88
43) Isopropyl Acetate	6.342	43	99901	28.506	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
Data File : VX042260.D
Acq On : 10 Jul 2024 12:26
Operator : JC/MD
Sample : PB161974ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161974ZHE#01

Quant Time: Jul 11 00:54:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

