

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042269.D
 Acq On : 10 Jul 2024 15:57
 Operator : JC/MD
 Sample : PB161974TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161974TB

Quant Time: Jul 11 00:58:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	163374	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	265421	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239345	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115183	61.140	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	122.280%#
35) Dibromofluoromethane	5.385	113	106847	56.734	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.460%
50) Toluene-d8	8.647	98	328564	49.661	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	119846	49.140	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.280%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	45850	55.426	ug/l	90
18) Methyl Acetate	2.715	43	4262	1.889	ug/l	97
20) Methylene Chloride	2.788	84	7892	4.357	ug/l #	85
43) Isopropyl Acetate	6.348	43	98937	24.742	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
Data File : VX042269.D
Acq On : 10 Jul 2024 15:57
Operator : JC/MD
Sample : PB161974TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161974TB

Quant Time: Jul 11 00:58:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

