

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042270.D
 Acq On : 10 Jul 2024 16:20
 Operator : JC/MD
 Sample : P3187-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3

Quant Time: Jul 11 00:59:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	170068	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	274917	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	121081	61.741	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 123.480%#		
35) Dibromofluoromethane	5.385	113	113610	58.241	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 116.480%		
50) Toluene-d8	8.647	98	358551	52.321	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 104.640%		
62) 4-Bromofluorobenzene	11.079	95	130107	51.505	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.000%		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	57596	66.885	ug/l	# 88
18) Methyl Acetate	2.709	43	3820	1.626	ug/l	# 90
20) Methylene Chloride	2.788	84	7665	4.065	ug/l	88
43) Isopropyl Acetate	6.348	43	95166	22.977	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
Data File : VX042270.D
Acq On : 10 Jul 2024 16:20
Operator : JC/MD
Sample : P3187-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3

Quant Time: Jul 11 00:59:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

