

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042271.D
 Acq On : 10 Jul 2024 16:43
 Operator : JC/MD
 Sample : P3183-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-226

Quant Time: Jul 11 00:59:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149597	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	243565	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108105	62.668	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 125.340%	#
35) Dibromofluoromethane	5.385	113	102282	59.183	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 118.360%	
50) Toluene-d8	8.647	98	316012	52.050	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.100%	
62) 4-Bromofluorobenzene	11.079	95	114319	51.080	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.160%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	57893	76.430	ug/l	# 87
18) Methyl Acetate	2.709	43	3782	1.830	ug/l	99
20) Methylene Chloride	2.788	84	8660	5.221	ug/l	84
25) 2-Butanone	4.593	43	6801	5.771	ug/l	97
43) Isopropyl Acetate	6.342	43	82590	22.507	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
Data File : VX042271.D
Acq On : 10 Jul 2024 16:43
Operator : JC/MD
Sample : P3183-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-226

Quant Time: Jul 11 00:59:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

