

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042276.D
 Acq On : 10 Jul 2024 18:40
 Operator : JC/MD
 Sample : P3187-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-1

Quant Time: Jul 11 01:02:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138019	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	233841	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91248	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	114807	72.136	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 144.280%#	
35) Dibromofluoromethane	5.385	113	108160	65.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 130.380%#	
50) Toluene-d8	8.647	98	308770	52.972	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.940%	
62) 4-Bromofluorobenzene	11.079	95	111660	51.966	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.940%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	60629	86.756	ug/l	# 86
18) Methyl Acetate	2.715	43	4331	2.272	ug/l	96
20) Methylene Chloride	2.788	84	7645	4.996	ug/l	# 87
25) 2-Butanone	4.587	43	6629	6.097	ug/l	# 87
43) Isopropyl Acetate	6.342	43	96771	27.468	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
Data File : VX042276.D
Acq On : 10 Jul 2024 18:40
Operator : JC/MD
Sample : P3187-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-1

Quant Time: Jul 11 01:02:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

