

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071024\  
 Data File : VX042277.D  
 Acq On : 10 Jul 2024 19:04  
 Operator : JC/MD  
 Sample : P3187-08  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 WC-2

Quant Time: Jul 11 01:03:02 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061824W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 21 23:51:58 2024  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550     | 168   | 193610   | 50.000   | ug/l       | # 0.00   |
| 34) 1,4-Difluorobenzene     | 6.763     | 114   | 301767   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055    | 117   | 284345   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 135535   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.952     | 65    | 120716   | 54.070   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 78 - 117 | Recovery | = 108.140% |          |
| 35) Dibromofluoromethane    | 5.385     | 113   | 116266   | 54.300   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 75 - 124 | Recovery | = 108.600% |          |
| 50) Toluene-d8              | 8.647     | 98    | 403637   | 53.660   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 92 - 112 | Recovery | = 107.320% |          |
| 62) 4-Bromofluorobenzene    | 11.079    | 95    | 146874   | 52.969   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 83 - 123 | Recovery | = 105.940% |          |
| Target Compounds            |           |       |          |          |            |          |
|                             |           |       |          |          | Qvalue     |          |
| 16) Acetone                 | 2.392     | 43    | 66616    | 67.953   | ug/l       | # 88     |
| 18) Methyl Acetate          | 2.715     | 43    | 4297     | 1.607    | ug/l       | 98       |
| 20) Methylene Chloride      | 2.788     | 84    | 7687     | 3.581    | ug/l       | # 84     |
| 25) 2-Butanone              | 4.586     | 43    | 8569     | 5.618    | ug/l       | # 77     |
| 43) Isopropyl Acetate       | 6.348     | 43    | 95243    | 20.949   | ug/l       | # 94     |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071024\  
Data File : VX042277.D  
Acq On : 10 Jul 2024 19:04  
Operator : JC/MD  
Sample : P3187-08  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
WC-2

Quant Time: Jul 11 01:03:02 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061824W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jun 21 23:51:58 2024  
Response via : Initial Calibration

