

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042279.D
 Acq On : 10 Jul 2024 19:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

Quant Time: Jul 11 01:04:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	152388	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	217860	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80876	46.025	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.040%
35) Dibromofluoromethane	5.385	113	79273	51.282	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.560%
50) Toluene-d8	8.647	98	259728	47.827	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.660%
62) 4-Bromofluorobenzene	11.079	95	104064	51.984	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.960%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	63739	54.902	ug/l	100
3) Chloromethane	1.300	50	68583	47.089	ug/l	99
4) Vinyl Chloride	1.374	62	69813	47.818	ug/l	99
5) Bromomethane	1.599	94	40292	55.889	ug/l	99
6) Chloroethane	1.666	64	36613	61.007	ug/l	97
7) Trichlorofluoromethane	1.874	101	108003	56.464	ug/l	98
8) Diethyl Ether	2.136	74	41171	45.854	ug/l	85
9) 1,1,2-Trichlorotrifluo...	2.319	101	67166	50.974	ug/l	# 84
10) Methyl Iodide	2.447	142	86578	52.865	ug/l	# 87
11) Tert butyl alcohol	2.995	59	87376	262.003	ug/l	# 93
12) 1,1-Dichloroethene	2.312	96	68059	51.781	ug/l	# 78
13) Acrolein	2.239	56	79381	212.346	ug/l	97
14) Allyl chloride	2.660	41	100872	48.392	ug/l	88
15) Acrylonitrile	3.068	53	220213	252.804	ug/l	98
16) Acetone	2.392	43	174257	225.839	ug/l	# 88
17) Carbon Disulfide	2.501	76	132133	49.304	ug/l	98
18) Methyl Acetate	2.709	43	97166	46.162	ug/l	92
19) Methyl tert-butyl Ether	3.117	73	224199	51.632	ug/l	99
20) Methylene Chloride	2.788	84	78256	46.318	ug/l	88
21) trans-1,2-Dichloroethene	3.087	96	68509	51.358	ug/l	85
22) Diisopropyl ether	3.770	45	220495	50.043	ug/l	# 81
23) Vinyl Acetate	3.721	43	979833	260.396	ug/l	# 94
24) 1,1-Dichloroethane	3.605	63	126374	51.217	ug/l	96
25) 2-Butanone	4.568	43	291709	243.006	ug/l	92
26) 2,2-Dichloropropane	4.471	77	82545	46.197	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	85452	51.147	ug/l	85
28) Bromochloromethane	4.897	49	49449	46.790	ug/l	# 67
29) Tetrahydrofuran	5.013	42	191850	247.462	ug/l	91
30) Chloroform	5.092	83	132179	51.419	ug/l	97
31) Cyclohexane	5.464	56	102361	49.810	ug/l	92
32) 1,1,1-Trichloroethane	5.379	97	114307	53.216	ug/l	95
36) 1,1-Dichloropropene	5.690	75	85516	52.096	ug/l	95
37) Ethyl Acetate	4.727	43	110253	51.560	ug/l	96
38) Carbon Tetrachloride	5.678	117	98168	54.126	ug/l	97
39) Methylcyclohexane	7.379	83	108026	52.036	ug/l	91
40) Benzene	6.037	78	287817	51.987	ug/l	98

07/11/2024
 Supervised By :Mahesh
 adoda

07/11/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
 Data File : VX042279.D
 Acq On : 10 Jul 2024 19:51
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

Quant Time: Jul 11 01:04:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	59244	51.110	ug/l	91
42) 1,2-Dichloroethane	6.092	62	92832	52.056	ug/l	91
43) Isopropyl Acetate	6.348	43	167949	51.169	ug/l #	93
44) Trichloroethene	7.123	130	80270	54.090	ug/l	84
45) 1,2-Dichloropropane	7.427	63	70939	52.111	ug/l	97
46) Dibromomethane	7.580	93	52173	52.601	ug/l #	73
47) Bromodichloromethane	7.824	83	97541	53.564	ug/l	95
48) Methyl methacrylate	7.696	41	82511	52.674	ug/l #	84
49) 1,4-Dioxane	7.665	88	41529	1001.920	ug/l #	80
51) 4-Methyl-2-Pentanone	8.574	43	576020	254.771	ug/l	93
52) Toluene	8.720	92	188418	53.446	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	96440	50.066	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	108791	54.567	ug/l #	84
55) 1,1,2-Trichloroethane	9.153	97	78354	53.351	ug/l	96
56) Ethyl methacrylate	9.116	69	123091	53.874	ug/l #	87
57) 1,3-Dichloropropane	9.311	76	123728	52.145	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	289397	271.804	ug/l	92
59) 2-Hexanone	9.433	43	446056	257.586	ug/l	91
60) Dibromochloromethane	9.519	129	88011	55.169	ug/l	100
61) 1,2-Dibromoethane	9.610	107	83340	54.718	ug/l	100
64) Tetrachloroethene	9.275	164	78167	53.001	ug/l	89
65) Chlorobenzene	10.079	112	220190	52.267	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	78767	53.403	ug/l	98
67) Ethyl Benzene	10.195	91	350909	53.151	ug/l	97
68) m/p-Xylenes	10.299	106	285609	107.166	ug/l	91
69) o-Xylene	10.640	106	142554	54.027	ug/l	91
70) Styrene	10.652	104	242662	54.844	ug/l	94
71) Bromoform	10.799	173	72615	50.962	ug/l #	98
73) Isopropylbenzene	10.963	105	361142	51.906	ug/l	96
74) N-amyl acetate	10.841	43	151805	49.119	ug/l	92
75) 1,1,2,2-Tetrachloroethane	11.213	83	125363	48.458	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	99822m	48.054	ug/l	
77) Bromobenzene	11.195	156	105662	50.991	ug/l	76
78) n-propylbenzene	11.305	91	393107	52.073	ug/l	96
79) 2-Chlorotoluene	11.366	91	244364	50.241	ug/l	91
80) 1,3,5-Trimethylbenzene	11.451	105	303377	52.350	ug/l	95
81) trans-1,4-Dichloro-2-b...	11.018	75	36144	49.413	ug/l	87
82) 4-Chlorotoluene	11.451	91	277349	51.583	ug/l	92
83) tert-Butylbenzene	11.713	119	324955	51.103	ug/l	89
84) 1,2,4-Trimethylbenzene	11.750	105	309326	53.203	ug/l	92
85) sec-Butylbenzene	11.890	105	379911	52.821	ug/l	95
86) p-Isopropyltoluene	12.006	119	335246	53.359	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	184193	50.571	ug/l	94
88) 1,4-Dichlorobenzene	12.042	146	187101	50.067	ug/l	95
89) n-Butylbenzene	12.335	91	250948	51.888	ug/l	97
90) Hexachloroethane	12.536	117	55884	52.245	ug/l	59
91) 1,2-Dichlorobenzene	12.335	146	191093	51.273	ug/l	94
92) 1,2-Dibromo-3-Chloropr...	12.945	75	26397	51.488	ug/l #	48
93) 1,2,4-Trichlorobenzene	13.585	180	128383	52.584	ug/l	95
94) Hexachlorobutadiene	13.725	225	62773	50.846	ug/l	99
95) Naphthalene	13.774	128	406007	49.497	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	133613	51.889	ug/l	94

07/11/2024
 Supervised By :Mahesh
 Dadoda

07/11/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071024\
Data File : VX042279.D
Acq On : 10 Jul 2024 19:51
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Quant Time: Jul 11 01:04:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED


Reviewed By :John
Carlone

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						


07/11/2024
Supervised By :Mahesh
Dadoda

07/11/2024



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 732
Delta R.T. -0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 15238
Ion Ratio Lower Upper
168 100
99 46.7 56.6 85.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#2
Dichlorodifluoromethane
Concen: 54.902 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 85 Resp: 63739
Ion Ratio Lower Upper
85 100
87 33.3 16.7 50.1







#3
Chloromethane
Concen: 47.089 ug/l
RT: 1.300 min Scan# 35
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 50 Resp: 6858
Ion Ratio Lower Upper
50 100
52 32.5 25.5 38.3



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#4
Vinyl Chloride
Concen: 47.818 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 62 Resp: 69813
Ion Ratio Lower Upper
62 100
64 32.3 25.4 38.0





#5
Bromomethane
Concen: 55.889 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 40292
Ion Ratio Lower Upper
94 100
96 92.2 74.6 111.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#6
Chloroethane
Concen: 61.007 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 64 Resp: 36613
Ion Ratio Lower Upper
64 100
66 33.2 25.1 37.7







#7
Trichlorofluoromethane
Concen: 56.464 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 10800
Ion Ratio Lower Upper
101 100
103 62.9 51.4 77.2



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#8
Diethyl Ether
Concen: 45.854 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 74 Resp: 41171
Ion Ratio Lower Upper
74 100
45 90.8 53.3 159.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.974 ug/l
RT: 2.319 min Scan# 201
Delta R.T. 0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 67160
Ion Ratio Lower Upper
101 100
85 39.9 37.0 55.6
151 90.0 58.6 87.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#10
Methyl Iodide
Concen: 52.865 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Tgt Ion:142 Resp: 86578
Ion Ratio Lower Upper
142 100
127 40.4 41.6 62.4#
141 14.1 11.6 17.4





#11

Tert butyl alcohol

Concen: 262.003 ug/l

RT: 2.995 min Scan# 31

Delta R.T. -0.012 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 87370

Ion Ratio Lower Upper

59 100



57 9.8 6.0 9.0

Manual Integrations

APPROVED

Reviewed By :John

Carlone





07/11/2024

Supervised By :Mahesh

Dadoda

07/11/2024



#12

1,1-Dichloroethene

Concen: 51.781 ug/l

RT: 2.312 min Scan# 201

Delta R.T. 0.006 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51




Tgt Ion: 96 Resp: 68059


Ion Ratio Lower Upper

96 100

61 139.5 144.6 216.8#


98 63.7 50.4 75.6





#13
Acrolein
Concen: 212.346 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 7938
Ion Ratio Lower Upper
56 100
55 69.3 57.1 85.7


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#14
Allyl chloride
Concen: 48.392 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 41 Resp: 100872
Ion Ratio Lower Upper
41 100
39 58.8 55.8 83.6
76 35.5 24.3 36.5





#15
Acrylonitrile
Concen: 252.804 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 22021
Ion Ratio Lower Upper
53 100
52 82.4 66.0 99.0
51 34.3 29.8 44.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#16
Acetone
Concen: 225.839 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 43 Resp: 174257
Ion Ratio Lower Upper
43 100
58 33.7 21.9 32.9#





#17
Carbon Disulfide
Concen: 49.304 ug/l
RT: 2.501 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 13213
Ion Ratio Lower Upper
76 100
78 9.4 7.0 10.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#18
Methyl Acetate
Concen: 46.162 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 43 Resp: 97166
Ion Ratio Lower Upper
43 100
74 25.9 17.8 26.8





#19

Methyl tert-butyl Ether

Concen: 51.632 ug/l

RT: 3.117 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 73 Resp: 224199

Ion Ratio Lower Upper

73 100



57 22.3 17.3 25.9

Manual Integrations

APPROVED

Reviewed By :John

Carlone





07/11/2024

Supervised By :Mahesh

Dadoda

07/11/2024



#20

Methylene Chloride

Concen: 46.318 ug/l

RT: 2.788 min Scan# 279

Delta R.T. 0.006 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Tgt Ion: 84 Resp: 78256




Ion Ratio Lower Upper


84 100

49 110.8 103.6 155.4

51 34.4 33.8 50.6


86 64.8 49.4 74.0





#21
trans-1,2-Dichloroethene
Concen: 51.358 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 68509
Ion Ratio Lower Upper
96 100
61 125.0 119.2 178.8
98 64.2 48.7 73.1


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#22
Diisopropyl ether
Concen: 50.043 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 45 Resp: 220495
Ion Ratio Lower Upper
45 100
43 64.8 69.7 104.5#
87 30.5 20.8 31.2
59 11.9 9.0 13.4





#23
 Vinyl Acetate
 Concen: 260.396 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX042279.D
 Acq: 10 Jul 2024 19:51



Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 43 Resp: 979833
 Ion Ratio Lower Upper
 43 100
 86 12.0 7.8 11.8


Manual Integrations
 APPROVED

Reviewed By :John
 Carlone




07/11/2024
 Supervised By :Mahesh
 Dadoda



07/11/2024




#24
 1,1-Dichloroethane
 Concen: 51.217 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX042279.D
 Acq: 10 Jul 2024 19:51



Tgt Ion: 63 Resp: 126374
 Ion Ratio Lower Upper
 63 100
 98 8.2 3.3 9.8
 100 5.0 2.1 6.5







#25
2-Butanone
Concen: 243.006 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 291709
Ion Ratio Lower Upper
43 100
72 28.5 19.4 29.2



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#26
2,2-Dichloropropane
Concen: 46.197 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 77 Resp: 82545
Ion Ratio Lower Upper
77 100
97 25.7 10.8 32.3






#27
cis-1,2-Dichloroethene
Concen: 51.147 ug/l
RT: 4.489 min Scan# 511
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :



VSTDCCC050EC



Tgt Ion: 96 Resp: 85452
Ion Ratio Lower Upper
96 100
61 129.1 0.0 312.2
98 65.1 0.0 129.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#28
Bromochloromethane
Concen: 46.790 ug/l
RT: 4.897 min Scan# 625
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51




Tgt Ion: 49 Resp: 49449
Ion Ratio Lower Upper
49 100
129 2.3 0.0 3.8
130 99.7 57.5 86.3#





#29
Tetrahydrofuran
Concen: 247.462 ug/l
RT: 5.013 min Scan# 644
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 42 Resp: 191850
Ion Ratio Lower Upper
42 100
72 48.4 34.2 51.4
71 45.3 32.2 48.2


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#30
Chloroform
Concen: 51.419 ug/l
RT: 5.092 min Scan# 657
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 83 Resp: 132179
Ion Ratio Lower Upper
83 100
85 65.0 50.3 75.5





#31
Cyclohexane
Concen: 49.810 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 10236
Ion Ratio Lower Upper
56 100
69 30.5 23.4 35.2
84 92.0 65.9 98.9


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda

07/11/2024



#32
1,1,1-Trichloroethane
Concen: 53.216 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 97 Resp: 114307
Ion Ratio Lower Upper
97 100
99 64.6 51.1 76.7
61 38.0 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 46.025 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 80870

Ion Ratio Lower Upper

65 100



67 53.7 0.0 100.2

Manual Integrations

APPROVED

Reviewed By :John

Carlone





07/11/2024

Supervised By :Mahesh

Dadoda

07/11/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51




Tgt Ion:114 Resp: 217860


Ion Ratio Lower Upper

114 100

63 17.4 0.0 45.0


88 14.9 0.0 36.6





#35
Dibromofluoromethane
Concen: 51.282 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 113 Resp: 7927
Ion Ratio Lower Upper
113 100
111 101.0 82.9 124.3
192 25.9 13.3 19.9


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#36
1,1-Dichloropropene
Concen: 52.096 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 75 Resp: 85516
Ion Ratio Lower Upper
75 100
110 38.8 17.0 50.9
77 31.1 24.7 37.1







#37
Ethyl Acetate
Concen: 51.560 ug/l
RT: 4.727 min Scan# 597
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 11025
Ion Ratio Lower Upper
43 100
61 15.1 10.6 16.0
70 11.8 8.3 12.5



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#38
Carbon Tetrachloride
Concen: 54.126 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion:117 Resp: 98168
Ion Ratio Lower Upper
117 100
119 94.7 78.2 117.2
121 31.3 24.4 36.6





#39

Methylcyclohexane

Concen: 52.036 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 108020

Ion Ratio Lower Upper

83 100

55 72.2 64.6 97.0



98 50.5 36.7 55.1

Manual Integrations

APPROVED

Reviewed By :John

Carlone





07/11/2024

Supervised By :Mahesh

Dadoda

07/11/2024



#40

Benzene

Concen: 51.987 ug/l

RT: 6.037 min Scan# 812

Delta R.T. 0.000 min

Lab File: VX042279.D




Acq: 10 Jul 2024 19:51


Tgt Ion: 78 Resp: 287817

Ion Ratio Lower Upper

78 100

77 23.2 19.3 28.9





#41
Methacrylonitrile
Concen: 51.110 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion: 41 Resp: 59244
Ion Ratio Lower Upper
41 100
39 53.8 45.2 67.8
67 78.2 52.6 78.8
52 31.3 22.7 34.1



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#42
1,2-Dichloroethane
Concen: 52.056 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 62 Resp: 92832
Ion Ratio Lower Upper
62 100
98 11.1 0.0 16.0





#43
Isopropyl Acetate
Concen: 51.169 ug/l
RT: 6.348 min Scan# 80
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
Client Sample Id :
VSTDCCC050EC



Tgt Ion: 43 Resp: 167949
Ion Ratio Lower Upper
43 100
61 22.1 15.2 22.8
87 15.2 9.8 14.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#44
Trichloroethene
Concen: 54.090 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion:130 Resp: 80270
Ion Ratio Lower Upper
130 100
95 86.4 0.0 204.4





#45

1,2-Dichloropropane

Concen: 52.111 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 70939

Ion Ratio Lower Upper

63 100



65 32.2 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :John

Carlone





07/11/2024

Supervised By :Mahesh

Dadoda

07/11/2024



#46

Dibromomethane

Concen: 52.601 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51




Tgt Ion: 93 Resp: 52173


Ion Ratio Lower Upper

93 100

95 85.5 66.6 100.0

174 133.5 69.4 104.0#







#47
Bromodichloromethane
Concen: 53.564 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 83 Resp: 9754
Ion Ratio Lower Upper
83 100
85 66.7 49.8 74.6
127 9.9 7.1 10.7



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#48
Methyl methacrylate
Concen: 52.674 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 41 Resp: 82511
Ion Ratio Lower Upper
41 100
69 90.7 59.2 88.8#
39 51.3 47.7 71.5





#49

1,4-Dioxane

Concen: 1001.920 ug/l

RT: 7.665 min Scan# 1080

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 41529
 Ion Ratio Lower Upper
 88 100
 43 24.0 0.0 0.0
 58 58.2 60.1 90.1



Manual Integrations
 APPROVED

Reviewed By :John
 Carlone



07/11/2024
 Supervised By :Mahesh
 Dadoda

07/11/2024



#50

Toluene-d8

Concen: 47.827 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Tgt Ion: 98 Resp: 259728
 Ion Ratio Lower Upper
 98 100
 100 68.1 53.1 79.7





#51
4-Methyl-2-Pentanone
Concen: 254.771 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X



ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 576020
Ion Ratio Lower Upper
43 100
58 42.1 30.4 45.6



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#52
Toluene
Concen: 53.446 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 92 Resp: 188418
Ion Ratio Lower Upper
92 100
91 169.2 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 50.066 ug/l

RT: 8.982 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 75 Resp: 96440

Ion Ratio Lower Upper

75 100



77 32.2 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :John

Carlone





07/11/2024

Supervised By :Mahesh

Dadoda

07/11/2024



#54

cis-1,3-Dichloropropene

Concen: 54.567 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51




Tgt Ion: 75 Resp: 108791


Ion Ratio Lower Upper

75 100

77 30.3 25.3 37.9

39 41.1 46.7 70.1#





#55
1,1,2-Trichloroethane
Concen: 53.351 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :




VSTDCCC050EC

Tgt Ion: 97	Resp: 78354
Ion Ratio	Lower Upper
97 100	
83 84.1	65.8 98.8
85 54.8	41.6 62.4
99 64.1	47.8 71.8

Manual Integrations


APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#56
Ethyl methacrylate
Concen: 53.874 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 69	Resp: 123091
Ion Ratio	Lower Upper
69 100	
41 61.3	56.8 85.2
39 29.8	32.0 48.0#







#57
1,3-Dichloropropane
Concen: 52.145 ug/l
RT: 9.311 min Scan# 1134
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 76 Resp: 123728
Ion Ratio Lower Upper
76 100
78 32.6 26.1 39.1



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#58
2-Chloroethyl Vinyl ether
Concen: 271.804 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion: 63 Resp: 289397
Ion Ratio Lower Upper
63 100
106 32.3 22.5 33.7





#59
2-Hexanone
Concen: 257.586 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 446050
Ion Ratio Lower Upper
43 100
58 58.6 26.2 78.5


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#60
Dibromochloromethane
Concen: 55.169 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion:129 Resp: 88011
Ion Ratio Lower Upper
129 100
127 77.8 38.8 116.4





#61
1,2-Dibromoethane
Concen: 54.718 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:107 Resp: 83340
Ion Ratio Lower Upper
107 100
109 93.6 75.2 112.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#62
4-Bromofluorobenzene
Concen: 51.984 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51




Tgt Ion: 95 Resp: 104064
Ion Ratio Lower Upper
95 100
174 99.5 0.0 135.6
176 96.6 0.0 131.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 202562
Ion Ratio Lower Upper
117 100
82 49.6 46.2 69.4
119 32.1 25.5 38.3


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#64
Tetrachloroethene
Concen: 53.001 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion:164 Resp: 78167
Ion Ratio Lower Upper
164 100
166 123.9 102.7 154.1
129 83.5 80.4 120.6
131 81.1 76.5 114.7





#65
Chlorobenzene
Concen: 52.267 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:112 Resp: 220190
Ion Ratio Lower Upper
112 100
114 32.6 25.7 38.5


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 53.403 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51




Tgt Ion:131 Resp: 78767
Ion Ratio Lower Upper
131 100
133 96.7 47.1 141.4
119 64.1 32.2 96.6





#67
Ethyl Benzene
Concen: 53.151 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 350909
Ion Ratio Lower Upper
91 100
106 32.7 24.9 37.3


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#68
m/p-Xylenes
Concen: 107.166 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion:106 Resp: 285609
Ion Ratio Lower Upper
106 100
91 192.9 165.5 248.3







#69
o-Xylene
Concen: 54.027 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 142554
Ion Ratio Lower Upper
106 100
91 203.0 108.9 326.8



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#70
Styrene
Concen: 54.844 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion:104 Resp: 242662
Ion Ratio Lower Upper
104 100
78 45.9 43.6 65.4
103 54.1 43.6 65.4





#71
Bromoform
Concen: 50.962 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:173 Resp: 72615
Ion Ratio Lower Upper
173 100
175 49.6 24.1 72.4
254 0.1 0.0 0.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51




Tgt Ion:152 Resp: 114427
Ion Ratio Lower Upper
152 100
115 73.5 43.3 129.9
150 172.6 0.0 345.0





#73
Isopropylbenzene
Concen: 51.906 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 361142
Ion Ratio Lower Upper
105 100
120 27.8 12.9 38.6


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#74
N-ethyl acetate
Concen: 49.119 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51




Tgt Ion: 43 Resp: 151805
Ion Ratio Lower Upper
43 100
70 48.1 32.9 49.3
55 25.8 19.3 28.9
61 26.5 18.8 28.2





#75
1,1,2,2-Tetrachloroethane
Concen: 48.458 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 12536
Ion Ratio Lower Upper
83 100
131 11.5 5.3 15.8
85 64.1 32.0 96.2


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#76
1,2,3-Trichloropropane
Concen: 48.054 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51




Tgt Ion: 75 Resp: 99822
Ion Ratio Lower Upper
75 100
77 43.4 19.1 57.5





#77
Bromobenzene
Concen: 50.991 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:156 Resp: 105662
Ion Ratio Lower Upper
156 100
77 127.4 88.4 265.3
158 98.4 48.0 144.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#78
n-propylbenzene
Concen: 52.073 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Tgt Ion: 91 Resp: 393107
Ion Ratio Lower Upper
91 100
120 25.0 11.5 34.4







#79
2-Chlorotoluene
Concen: 50.241 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 244364
Ion Ratio Lower Upper
91 100
126 37.3 16.3 48.8



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#80
1,3,5-Trimethylbenzene
Concen: 52.350 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion:105 Resp: 303377
Ion Ratio Lower Upper
105 100
120 51.6 24.1 72.3







#81
trans-1,4-Dichloro-2-butene
Concen: 49.413 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 75 Resp: 36144
Ion Ratio Lower Upper
75 100
53 92.9 90.5 135.7
89 46.4 36.9 55.3



Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#82
4-Chlorotoluene
Concen: 51.583 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion: 91 Resp: 277349
Ion Ratio Lower Upper
91 100
126 32.8 14.2 42.6





#83

tert-Butylbenzene

Concen: 51.103 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion:119 Resp: 32495

Ion	Ratio	Lower	Upper
119	100		
91	52.1	31.2	93.6
134	24.9	11.7	35.0

Manual Integrations



APPROVED

Reviewed By :John
Carlone



07/11/2024
Supervised By :Mahesh
Dadoda

07/11/2024



#84

1,2,4-Trimethylbenzene

Concen: 53.203 ug/l

RT: 11.750 min Scan# 1749




Delta R.T. 0.000 min


Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Tgt Ion:105 Resp: 309326

Ion	Ratio	Lower	Upper
105	100		
120	47.3	21.2	63.6







#85
 sec-Butylbenzene
 Concen: 52.821 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX042279.D
 Acq: 10 Jul 2024 19:51

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion:105 Resp: 37991
 Ion Ratio Lower Upper
 105 100
 134 22.0 9.8 29.4



Manual Integrations
 APPROVED

Reviewed By :John
 Carlone






07/11/2024
 Supervised By :Mahesh
 Dadoda


07/11/2024



#86
 p-Isopropyltoluene
 Concen: 53.359 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX042279.D
 Acq: 10 Jul 2024 19:51

Tgt Ion:119 Resp: 335246
 Ion Ratio Lower Upper
 119 100
 134 26.7 12.8 38.3
 91 21.1 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 50.571 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :



VSTDCCC050EC

Tgt Ion:146 Resp: 18419

Ion	Ratio	Lower	Upper
146	100		
111	36.7	22.1	66.1
148	65.4	31.8	95.3



Manual Integrations
APPROVED

Reviewed By :John
Carlone



07/11/2024
Supervised By :Mahesh
Dadoda

07/11/2024



#88

1,4-Dichlorobenzene

Concen: 50.067 ug/l

RT: 12.042 min Scan# 1797




Delta R.T. 0.000 min


Lab File: VX042279.D

Acq: 10 Jul 2024 19:51

Tgt Ion:146 Resp: 187101


Ion	Ratio	Lower	Upper
146	100		
111	35.3	21.3	63.7
148	64.4	31.8	95.4





#89
n-Butylbenzene
Concen: 51.888 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 250948
Ion Ratio Lower Upper
91 100
92 55.1 27.4 82.0
134 30.1 12.6 37.8


Manual Integrations
APPROVED

Reviewed By :John
Carlone






07/11/2024
Supervised By :Mahesh
Dadoda


07/11/2024



#90
Hexachloroethane
Concen: 52.245 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51


Tgt Ion:117 Resp: 55884
Ion Ratio Lower Upper
117 100
201 98.1 32.9 98.6





#91
1,2-Dichlorobenzene
Concen: 51.273 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 191093
Ion Ratio Lower Upper
146 100
111 37.3 22.4 67.0
148 64.8 31.6 95.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 51.488 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51




Tgt Ion: 75 Resp: 26397
Ion Ratio Lower Upper
75 100
155 111.1 36.1 108.5#
157 143.4 44.9 134.7#





#93
1,2,4-Trichlorobenzene
Concen: 52.584 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:180 Resp: 12838
Ion Ratio Lower Upper
180 100
182 94.8 48.3 144.8
145 27.6 18.0 54.0


Manual Integrations
APPROVED

Reviewed By :John
Carlone




07/11/2024
Supervised By :Mahesh
Dadoda



07/11/2024




#94
Hexachlorobutadiene
Concen: 50.846 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51



Tgt Ion:225 Resp: 62773
Ion Ratio Lower Upper
225 100
223 63.0 31.1 93.2
227 64.4 32.5 97.5






#95
Naphthalene
Concen: 49.497 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Instrument :

MSVOA_X

ClientSampleId :



VSTDCCC050EC



Tgt Ion:128 Resp: 40600
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.9 8.8 13.2


Manual Integrations
APPROVED

Reviewed By :John
Carlone



07/11/2024
Supervised By :Mahesh
Dadoda

07/11/2024



#96
1,2,3-Trichlorobenzene
Concen: 51.889 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX042279.D
Acq: 10 Jul 2024 19:51

Tgt Ion:180 Resp: 133613
Ion Ratio Lower Upper
180 100
182 97.5 46.9 140.6
145 29.1 17.8 53.5

