

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
 Data File : VX046935.D
 Acq On : 10 Jul 2025 11:44
 Operator : JC/MD
 Sample : VX0710WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0710WBL01

Quant Time: Jul 11 01:26:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	311309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	568946	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	543608	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	271904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	232769	56.598	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.200%	
35) Dibromofluoromethane	5.391	113	191587	49.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.220%	
50) Toluene-d8	8.646	98	692184	50.799	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.600%	
62) 4-Bromofluorobenzene	11.079	95	266771	51.513	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.020%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
Data File : VX046935.D
Acq On : 10 Jul 2025 11:44
Operator : JC/MD
Sample : VX0710WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0710WBL01

Quant Time: Jul 11 01:26:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

