

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
 Data File : VX046939.D
 Acq On : 10 Jul 2025 13:18
 Operator : JC/MD
 Sample : Q2553-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Jul 11 01:29:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	354436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	642674	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	599480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	308124	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	255734	54.615	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.240%
35) Dibromofluoromethane	5.397	113	213280	48.890	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.780%
50) Toluene-d8	8.653	98	772213	50.170	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.340%
62) 4-Bromofluorobenzene	11.079	95	302008	51.627	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
Data File : VX046939.D
Acq On : 10 Jul 2025 13:18
Operator : JC/MD
Sample : Q2553-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Jul 11 01:29:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

