

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
 Data File : VX046940.D
 Acq On : 10 Jul 2025 13:39
 Operator : JC/MD
 Sample : Q2553-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Jul 11 01:29:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	355069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	638277	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	597675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	300528	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	255926	54.559	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.120%	
35) Dibromofluoromethane	5.397	113	213797	49.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.700%	
50) Toluene-d8	8.653	98	767737	50.223	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	11.079	95	296852	51.096	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.200%	
Target Compounds						
17) Carbon Disulfide	2.532	76	13937	1.380	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
Data File : VX046940.D
Acq On : 10 Jul 2025 13:39
Operator : JC/MD
Sample : Q2553-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Jul 11 01:29:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

