

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
 Data File : VX046942.D
 Acq On : 10 Jul 2025 14:21
 Operator : JC/MD
 Sample : Q2554-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 250709059-11 Trip blank

Quant Time: Jul 11 01:30:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	339396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	600517	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	555080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	279121	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	244408	54.510	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.020%
35) Dibromofluoromethane	5.397	113	201406	49.409	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.820%
50) Toluene-d8	8.653	98	722653	50.247	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.500%
62) 4-Bromofluorobenzene	11.079	95	275498	50.402	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
Data File : VX046942.D
Acq On : 10 Jul 2025 14:21
Operator : JC/MD
Sample : Q2554-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
250709059-11 Trip blank

Quant Time: Jul 11 01:30:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

