

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
 Data File : VX046952.D
 Acq On : 10 Jul 2025 17:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 11 01:35:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	468309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	823051	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	760993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	384583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	319102	51.578	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.160%		
35) Dibromofluoromethane	5.397	113	269984	48.325	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.640%		
50) Toluene-d8	8.653	98	989697	50.208	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.420%		
62) 4-Bromofluorobenzene	11.079	95	380518	50.793	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.580%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071025\
Data File : VX046952.D
Acq On : 10 Jul 2025 17:55
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 11 01:35:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

