

Data File : VX003325.D
Acq On : 11 Jul 2018 17:42
Operator : JC/MD
Sample : J3934-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TS-BASELINE-WATER

Quant Time: Oct 12 05:49:01 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	255404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	388910	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	361814	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	197475	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174880	43.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.02%	
35) Dibromofluoromethane	5.51	113	147162	35.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.44%	
50) Toluene-d8	8.72	98	559862	37.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	75.44%	
62) 4-Bromofluorobenzene	11.14	95	190203	34.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	69.36%	
Target Compounds						
16) Acetone	2.45	43	18158	12.724	ug/l	93
17) Carbon Disulfide	2.57	76	15248	1.912	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX003325.D
Acq On : 11 Jul 2018 17:42
Operator : JC/MD
Sample : J3934-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TS-BASELINE-WATER

Quant Time: Oct 12 05:49:01 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

