

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071118\
 Data File : VX003328.D
 Acq On : 11 Jul 2018 19:02
 Operator : JC/MD
 Sample : PB110983ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110983ZHE#02

Quant Time: Jul 12 04:34:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	292052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	435874	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	410899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	223017	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	190681	41.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.98%	
35) Dibromofluoromethane	5.51	113	164941	35.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.44%	
50) Toluene-d8	8.72	98	625352	37.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.20%	
62) 4-Bromofluorobenzene	11.14	95	214773	34.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.90%	
Target Compounds						
10) Methyl Iodide	2.52	142	3303	6.387	ug/l	95
16) Acetone	2.45	43	20576	12.610	ug/l #	88
18) Methyl Acetate	2.78	43	27171	5.902	ug/l	95
20) Methylene Chloride	2.86	84	5706	1.015	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071118\
Data File : VX003328.D
Acq On : 11 Jul 2018 19:02
Operator : JC/MD
Sample : PB110983ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB110983ZHE#02

Quant Time: Jul 12 04:34:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 09 15:24:16 2018
Response via : Initial Calibration

