

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071118\
 Data File : VX003332.D
 Acq On : 11 Jul 2018 20:49
 Operator : JC/MD
 Sample : PB110983ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB110983ZHE#06

Quant Time: Jul 12 04:36:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X070918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 09 15:24:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	278202	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	414321	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	353940	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	192659	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182562	41.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.40%	
35) Dibromofluoromethane	5.51	113	157895	35.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.94%	
50) Toluene-d8	8.72	98	546882	34.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.18%	
62) 4-Bromofluorobenzene	11.14	95	186629	31.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.90%	
Target Compounds						
					Qvalue	
10) Methyl Iodide	2.52	142	2042	6.128	ug/l #	92
16) Acetone	2.45	43	22584	14.529	ug/l #	85
18) Methyl Acetate	2.78	43	28606	6.524	ug/l	95
20) Methylene Chloride	2.86	84	5707	1.101	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

