

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071119\
 Data File : VX010801.D
 Acq On : 11 Jul 2019 12:53
 Operator : JC/SP
 Sample : K3726-03ME
 Misc : 5.87g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PX-3_070919ME

Quant Time: Jul 12 13:16:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	206429	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.85	114	316382	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	282277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148979	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	101908	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
35) Dibromofluoromethane	5.50	113	91832	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
50) Toluene-d8	8.71	98	371107	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
62) 4-Bromofluorobenzene	11.14	95	135226	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
Target Compounds						
68) m/p-Xylenes	10.36	106	4531	1.209	ug/l	75
69) o-Xylene	10.70	106	8503	2.306	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	142437	17.109	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	146712	17.540	ug/l	100
85) sec-Butylbenzene	11.94	105	19304	1.968	ug/l	91
86) p-Isopropyltoluene	12.07	119	52688	5.806	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071119\
Data File : VX010801.D
Acq On : 11 Jul 2019 12:53
Operator : JC/SP
Sample : K3726-03ME
Misc : 5.87g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PX-3_070919ME

Quant Time: Jul 12 13:16:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

