

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071119\
 Data File : VX010805.D
 Acq On : 11 Jul 2019 14:27
 Operator : JC/SP
 Sample : K3749-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB071019

Quant Time: Jul 12 13:09:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235684	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	367371	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	340045	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156856	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120391	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
35) Dibromofluoromethane	5.50	113	111176	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	8.71	98	440576	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
62) 4-Bromofluorobenzene	11.13	95	150388	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
Target Compounds						
16) Acetone	2.45	43	5659	5.110	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071119\
Data File : VX010805.D
Acq On : 11 Jul 2019 14:27
Operator : JC/SP
Sample : K3749-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB071019

Quant Time: Jul 12 13:09:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

