

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071119\
 Data File : VX010806.D
 Acq On : 11 Jul 2019 14:50
 Operator : JC/SP
 Sample : K3757-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-02-190710

Quant Time: Jul 12 07:23:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	228114	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	359924	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	332588	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149045	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117316	54.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.70%	
35) Dibromofluoromethane	5.50	113	107860	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	8.71	98	426099	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	11.13	95	145089	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
Target Compounds						
16) Acetone	2.45	43	6362	5.935	ug/l	93
20) Methylene Chloride	2.86	84	33563	15.215	ug/l	98
52) Toluene	8.78	92	140448	23.934	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071119\
Data File : VX010806.D
Acq On : 11 Jul 2019 14:50
Operator : JC/SP
Sample : K3757-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-02-190710

Quant Time: Jul 12 07:23:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

