

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071119\
 Data File : VX010807.D
 Acq On : 11 Jul 2019 15:13
 Operator : JC/SP
 Sample : K3756-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LC-TB02-190710

Quant Time: Jul 12 07:24:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	236477	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	368073	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	335258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	154853	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122061	55.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.10%	
35) Dibromofluoromethane	5.50	113	113618	50.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.90%	
50) Toluene-d8	8.71	98	446018	51.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.54%	
62) 4-Bromofluorobenzene	11.13	95	151338	48.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071119\
Data File : VX010807.D
Acq On : 11 Jul 2019 15:13
Operator : JC/SP
Sample : K3756-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LC-TB02-190710

Quant Time: Jul 12 07:24:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

