

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071119\
 Data File : VX010816.D
 Acq On : 11 Jul 2019 18:44
 Operator : JC/SP
 Sample : K3749-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jul 12 08:07:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	215751	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	338441	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	312354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	143991	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110086	54.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.84%	
35) Dibromofluoromethane	5.50	113	103572	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
50) Toluene-d8	8.71	98	403605	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	11.13	95	137429	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
Target Compounds						
16) Acetone	2.45	43	4965	4.897	ug/l	95
27) cis-1,2-Dichloroethene	4.59	96	791	0.340	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071119\
Data File : VX010816.D
Acq On : 11 Jul 2019 18:44
Operator : JC/SP
Sample : K3749-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jul 12 08:07:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

