

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071122\
 Data File : VX029996.D
 Acq On : 11 Jul 2022 14:46
 Operator : JC/MD
 Sample : PB146125TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB146125TB

Quant Time: Jul 12 02:02:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	189336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	315492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96718	40.355	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.720%
35) Dibromofluoromethane	5.385	113	92752	47.043	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.080%
50) Toluene-d8	8.647	98	355921	48.119	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.240%
62) 4-Bromofluorobenzene	11.079	95	116640	44.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.220%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18113	15.448	ug/l	97
20) Methylene Chloride	2.788	84	11675	4.715	ug/l	96
37) Ethyl Acetate	4.721	43	16096	4.488	ug/l #	95
43) Isopropyl Acetate	6.342	43	67959	12.583	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071122\
Data File : VX029996.D
Acq On : 11 Jul 2022 14:46
Operator : JC/MD
Sample : PB146125TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB146125TB

Quant Time: Jul 12 02:02:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

