

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071122\
 Data File : VX029997.D
 Acq On : 11 Jul 2022 15:09
 Operator : JC/MD
 Sample : PB146107#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB146107#01

Quant Time: Jul 12 02:03:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	194745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	332639	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288607	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	101807	41.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.600%
35) Dibromofluoromethane	5.391	113	97915	47.101	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.653	98	375194	48.110	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.220%
62) 4-Bromofluorobenzene	11.085	95	123962	44.967	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.940%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	17936	14.872	ug/l	92
20) Methylene Chloride	2.794	84	10904	4.281	ug/l	94
37) Ethyl Acetate	4.721	43	16648	4.403	ug/l	99
43) Isopropyl Acetate	6.348	43	69229	12.158	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071122\
Data File : VX029997.D
Acq On : 11 Jul 2022 15:09
Operator : JC/MD
Sample : PB146107#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB146107#01

Quant Time: Jul 12 02:03:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

