

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071122\
 Data File : VX030006.D
 Acq On : 11 Jul 2022 18:41
 Operator : JC/MD
 Sample : N3648-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CMR-TP2-0708

Quant Time: Jul 12 02:05:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	189981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	327417	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103257	42.937	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.880%
35) Dibromofluoromethane	5.391	113	96972	47.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	8.652	98	371094	48.343	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.680%
62) 4-Bromofluorobenzene	11.085	95	125348	46.195	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.400%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	33736	28.674	ug/l	98
20) Methylene Chloride	2.788	84	23295	9.375	ug/l	86
25) 2-Butanone	4.562	43	10684	5.232	ug/l	100
37) Ethyl Acetate	4.720	43	37200	9.994	ug/l	98
40) Benzene	6.043	78	16195	1.840	ug/l	93
43) Isopropyl Acetate	6.348	43	144950	25.861	ug/l	99
52) Toluene	8.726	92	9662	1.820	ug/l	97
70) Styrene	10.658	104	10377	1.617	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071122\
Data File : VX030006.D
Acq On : 11 Jul 2022 18:41
Operator : JC/MD
Sample : N3648-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CMR-TP2-0708

Quant Time: Jul 12 02:05:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

