

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071218\
 Data File : VX003349.D
 Acq On : 12 Jul 2018 14:07
 Operator : JC/MD
 Sample : VX0712WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0712WBL01

Quant Time: Jul 12 17:09:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X071218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 12 13:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	62432	30.00	ug/l	-0.01
28) 1,4-Difluorobenzene	6.86	114	309356	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	277887	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	128684	30.18	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.60%
60) 4-Bromofluorobenzene	11.14	95	121306	27.17	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.57%
63) Toluene-d8	8.72	98	363945	29.62	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.73%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071218\
Data File : VX003349.D
Acq On : 12 Jul 2018 14:07
Operator : JC/MD
Sample : VX0712WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0712WBL01

Quant Time: Jul 12 17:09:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X071218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 12 13:28:20 2018
Response via : Initial Calibration

