

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071218\
 Data File : VX003357.D
 Acq On : 12 Jul 2018 17:48
 Operator : JC/MD
 Sample : J3952-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EFF-WW

Quant Time: Jul 12 19:15:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X071218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 12 13:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	60623	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	316337	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	290749	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	123080	29.73	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.10%
60) 4-Bromofluorobenzene	11.14	95	136395	29.19	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	97.30%
63) Toluene-d8	8.72	98	376156	29.26	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.53%

Target Compounds

					Ovalue
15) Acetone	2.45	58	136175	225.457	ug/l 77
25) Chloroform	5.22	83	19863	2.902	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071218\
Data File : VX003357.D
Acq On : 12 Jul 2018 17:48
Operator : JC/MD
Sample : J3952-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EFF-WW

Quant Time: Jul 12 19:15:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X071218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 12 13:28:20 2018
Response via : Initial Calibration

