

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071218\
 Data File : VX003349.D
 Acq On : 12 Jul 2018 14:07
 Operator : JC/MD
 Sample : VX0712WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0712WBL01

Quant Time: Jul 12 17:09:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X071218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 12 13:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	62432	30.00	ug/l	-0.01
28) 1,4-Difluorobenzene	6.86	114	309356	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	277887	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	128684	30.18	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.60%
60) 4-Bromofluorobenzene	11.14	95	121306	27.17	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.57%
63) Toluene-d8	8.72	98	363945	29.62	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.73%


Target Compounds	Ovalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071218\
Data File : VX003349.D
Acq On : 12 Jul 2018 14:07
Operator : JC/MD
Sample : VX0712WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0712WBL01

Quant Time: Jul 12 17:09:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X071218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 12 13:28:20 2018
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 30.000 ug/l

RT: 5.02 min Scan# 727

Delta R.T. -0.01 min

Lab File: VX003349.D

Acq: 12 Jul 2018 14:07

Instrument :

MSVOA_X

ClientSampleId :

VX0712WBL01


Tot Ion:128 Resp: 62432

Ion Ratio Lower Upper

128 100

49 152.6 0.0 423.0




130 130.1 0.0 324.3



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 49.00 (48.70 to 49.70): VX0

Ion 130.00 (129.70 to 130.70): V



#27

1,2-Dichloroethane-d4

Concen: 30.180 ug/l

RT: 6.07 min Scan# 900

Delta R.T. -0.00 min

Lab File: VX003349.D


Acq: 12 Jul 2018 14:07

Tgt Ion: 65 Resp: 128684

Ion Ratio Lower Upper



65 100


67 54.1 40.9 61.3



Abundance Ion 65.00 (64.70 to 65.70): VX0

Ion 67.00 (66.70 to 67.70): VX0









#28
 1,4-Difluorobenzene
 Concen: 30.000 ug/l
 RT: 6.86 min Scan# 1029
 Delta R.T. -0.01 min
 Lab File: VX003349.D
 Acq: 12 Jul 2018 14:07

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0712WBL01

Tot Ion:114 Resp: 309356
 Ion Ratio Lower Upper
 114 100
 63 19.8 17.1 25.7
 88 15.3 12.6 19.0




Abundance Ion 114.00 (113.70 to 114.70): V
 Ion 63.00 (62.70 to 63.70): VX
 Ion 88.00 (87.70 to 88.70): VX






#57
 Chlorobenzene-d5
 Concen: 30.000 ug/l
 RT: 10.12 min Scan# 1563
 Delta R.T. -0.00 min
 Lab File: VX003349.D
 Acq: 12 Jul 2018 14:07

Tgt Ion:117 Resp: 277887
 Ion Ratio Lower Upper
 117 100
 82 53.2 44.3 66.5
 119 31.9 25.7 38.5



Abundance Ion 117.00 (116.70 to 117.70): V
 Ion 82.00 (81.70 to 82.70): VX
 Ion 119.00 (118.70 to 119.70): V





#60

4-Bromofluorobenzene

Concen: 27.166 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.00 min

Lab File: VX003349.D

Acq: 12 Jul 2018 14:07

Instrument :

MSVOA_X

ClientSampleId :

VX0712WBL01


Tot Ion: 95 Resp: 121306

Ion Ratio Lower Upper

95 100

174 92.3 71.8 107.6




176 90.2 70.1 105.1



Abundance Ion 95.00 (94.70 to 95.70): VX0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V



#63

Toluene-d8

Concen: 29.625 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. -0.00 min

Lab File: VX003349.D


Acq: 12 Jul 2018 14:07

Tgt Ion: 98 Resp: 363945

Ion Ratio Lower Upper

98 100

100 64.0 50.8 76.2



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

