

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071219\
 Data File : VX010824.D
 Acq On : 12 Jul 2019 10:24
 Operator : JC/SP
 Sample : VX0712WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0712WBL01

Quant Time: Jul 13 01:54:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	206465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	322998	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	290887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126644	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107743	55.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.32%	
35) Dibromofluoromethane	5.49	113	101802	51.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.02%	
50) Toluene-d8	8.71	98	384471	50.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.70%	
62) 4-Bromofluorobenzene	11.13	95	122140	44.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071219\
Data File : VX010824.D
Acq On : 12 Jul 2019 10:24
Operator : JC/SP
Sample : VX0712WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0712WBL01

Quant Time: Jul 13 01:54:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

