

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071219\
 Data File : VX010828.D
 Acq On : 12 Jul 2019 12:09
 Operator : JC/SP
 Sample : K3782-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LC-TB03-190711

Quant Time: Jul 13 02:10:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	353677	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	321107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146234	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	115348	54.40	ug/l	0.00
Spiked Amount						
			Recovery	=	108.80%	
35) Dibromofluoromethane	5.50	113	108966	50.36	ug/l	0.00
Spiked Amount						
			Recovery	=	100.72%	
50) Toluene-d8	8.71	98	424584	51.29	ug/l	0.00
Spiked Amount						
			Recovery	=	102.58%	
62) 4-Bromofluorobenzene	11.13	95	143829	48.12	ug/l	0.00
Spiked Amount						
			Recovery	=	96.24%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071219\
Data File : VX010828.D
Acq On : 12 Jul 2019 12:09
Operator : JC/SP
Sample : K3782-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LC-TB03-190711

Quant Time: Jul 13 02:10:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

