

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071219\
 Data File : VX010831.D
 Acq On : 12 Jul 2019 13:19
 Operator : JC/SP
 Sample : K3782-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030LC011-SW-190711

Quant Time: Jul 13 02:17:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	206484	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	330053	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	303003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137940	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	107794	55.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.36%	
35) Dibromofluoromethane	5.49	113	102438	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
50) Toluene-d8	8.71	98	395280	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	11.13	95	132868	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
Target Compounds						
16) Acetone	2.45	43	4051	4.175	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071219\
Data File : VX010831.D
Acq On : 12 Jul 2019 13:19
Operator : JC/SP
Sample : K3782-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY030LC011-SW-190711

Quant Time: Jul 13 02:17:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

